CZECH METROLOGY INSTITUTE

Český metrologický institut, Úsek legální metrologie – Referát MPZ

Z 7002

CERTIFICATE

OF PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A0130-16

Subject of the ILC:

WORKING LENGTH GAUGES

Carried out in the period of: April 2016 – April 2017

ILC identification: **11-0100-16**

Participant:

S.C. METROMAT S.R.L.

Str. Ady Endre nr. 44, 505600 Sacele, jud. Brasov, Romania

Final statement

Following the Report No. 0318-ZV-A0100-16 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **18.05.2017**

Český metrologický institut

-4-

RNDr. Simona Klenovská
Head of the ILC Department

Ing. František Staněk, PhD.
Director for Legal Metrology

33 de laboratoare participante	**Valori E_n obtinute de Metromat**
Cehia (25) , Slovacia (4), Croatia (1), Italia (2), Israel (1), **Romania (Metromat)**	Micrometru: E_n = 0,28; 0,35; 0,11; -0,07; 0,48; -0,04; 0,06; -0,06; 0,06; -0,20; **Media valorilor E_n: 0,171** Subler : E_n= 0,00; 0,00; -0,35; -0.35; 0,00; **Media valorilor E_n: 0,140** Calibru neted: E_n= 0,00; -0,13; -0,32; -0,21; **Media valorilor E_n: 0,165** Cale de reglare: E_n= - 0,59; -0,74; **Media valorilor E_n: 0,665**

CZECH METROLOGY INSTITUTE

Český metrologický institut, Úsek legální metrologie – referát MPZ

Z 7002

CERTIFICATE

ON PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A0315-15

Subject of the ILC:

GAUGE BLOCKS

Carried out in the period of: March – December 2015

Identification of the ILC: **ILC 11-0300-15**

Participant:

METROMAT SRL

Str. Ady Endre no. 44, 505600 Sacele - Brasov, Romania

Final statement

Following the Report No. 0318-ZV-A0300-15 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **29.01.2016**

Český metrologický institut

-4-

RNDr. Simona K l e n o v s k á
Head of the Department for ILC

Ing. František S t a n ě k, PhD.
Director for Legal Metrology

17 Laboratoare participante	Valori E_n obtinute de Metromat
Cehia (12); Slovacia, Polonia, Israel , Sri Lanka, **Romania (Metromat)**	E_n = -0,461; -0,264; -0,443; -0,134; -0,060; -0,278; 0,447; 0,179; **Media valorilor E_n : 0,283**

CZECH METROLOGY INSTITUTE

Český metrologický institut, Úsek legální metrologie – referát MPZ

Z 7002

CERTIFICATE

ON PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A90110-13

Subject of the ILC:

CYLINDRICAL GAUGES AND SETTING GAUGES FOR MICROMETERS

Carried out in the period of: April – September 2013

Identification of the ILC: **ILC 11-100-13**

Participant:

METROMAT SRL

P.ta. Libertatii No. 2, 505600 Sacele, jud. Brasov, Romania

Final statement

Following the Report No. 0318-ZV-A90100-13 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **20.09.2013**

Český metrologický institut

-4-

RNDr. Simona K l e n o v s k á
Head of the Department for ILC

Ing. Jindřich P o š v á ř
Director for Legal Metrology

13 Laboratoare participante	Valori E_n obtinute de Metromat
Cehia (12),	E_n = 0,749; 0,887; 0,693; 0,499; 0,026; -0,205; 0,179; 0,282; 0,253; -0,214
Romania (Metromat)	**Media valorilor E_n: 0,398**

ROLAB - ASOCIATIA LABORATOARELOR DIN ROMANIA
Bucuresti, B-dul Unirii, nr. 61, bl. F3, sc. TR2, et. 3, ap. 301, sector 3
office_rolab@yahoo.com, fax. 031.817.45.24

Quality is a state of mind

ACREDITAT RENAR SR EN ISO/CEI 17043.2010
Certificat de acreditare Nr. IC 001/28.05.2012

CERTIFICAT MRC ISO 9001
Certificat Nr. 279/2010

acreditat pentru
FURNIZARE DE ÎNCERCĂRI
DE COMPETENȚĂ
RENAR
SR EN ISO/CEI 17043:2010
CERTIFICAT DE ACREDITARE
nr. IC 001/2012

CERTIFICAT

Nr. *250*

Prin prezentul document, se certifica participarea laboratorului apartinand ***SC METROMAT SRL - Sacele, jud. Brasov*** *la incercarile de competenta privind evaluarea performantelor laboratoarelor de etalonari sau incercari a mijloacelor de masurare:* ***comparatoare mecanice cu cadran sau comparatoare cu afişare digital.***
Rezultatele laboratorului sunt prezentate in ***Raportul incercarii de competenta nr. 06/2014 – ROLAB****, emis la data* ***de 02.06.2014****.*

Codul laboratorului Dv. este: ***104.***

Director ROLAB-IC

ing. Roberta TODOR

ASOCIATIA LABORATOARELOR DIN ROMANIA * ROLAB

Coordonator schema IC:
Etalonari Lungimi

dr.ing. Alexandru DUTA

7 Laboratoare participante	Valori E_n obtinute de Metromat
Din Romania	E_n = -0,11; -0,11; 0,00; -0,22; 0,00; -0,21; -0,11; 0,00; 0,11; 0,53 ; **Media valorilor E_n: 0,140**

Quality is a state of mind

ROLAB - ASOCIATIA LABORATOARELOR DIN ROMANIA
Bucuresti, B-dul Unirii, nr. 61, bl. F3, sc. TR2, et. 3, ap. 301, sector 3
office_rolab@yahoo.com, fax. 031.817.45.24

ACREDITAT RENAR SR EN ISO/CEI 17043.2010
Certificat de acreditare Nr. IC 001/28.05.2012

CERTIFICAT MRC ISO 9001
Certificat Nr. 279/2010

acreditat pentru
FURNIZARE DE ÎNCERCĂRI DE COMPETENȚĂ
SR EN ISO/CEI 17043:2010
CERTIFICAT DE ACREDITARE
nr. IC 001/2012

CERTIFICAT

Nr. *251*

Prin prezentul document, se certifica participarea laboratorului apartinand ***SC METROMAT SRL - Sacele, jud. Brasov*** *la incercarile de competenta privind evaluarea performantelor laboratoarelor de etalonari sau incercari a mijloacelor de masurare:* ***şublere mecanice cu vernier sau şublere cu afişare digitală.***

Rezultatele laboratorului sunt prezentate in ***Raportul incercarii de competenta nr. 05/2014 – ROLAB****, emis la data* ***de 22.05.2014****.*

Codul laboratorului Dv. este: ***104.***

Director ROLAB-IC

ing. Roberta TODOR

ASOCIATIA LABORATOARELOR DIN ROMANIA * ROLAB

Coordonator schema IC:
Etalonari Lungimi

dr.ing. Alexandru DUTA

16 Laboratoare participante	Valori E_n obtinute de Metromat
Din Romania	E_n = 0,21; 0,21; 0,43; 0,11; 0,11; 0,21; 0,32; **Media valorilor E_n: 0,228**

CZECH METROLOGY INSTITUTE

Český metrologický institut, Referát MPZ

Z 7002

CERTIFICATE

OF PARTICIPATION IN THE INTERLABORATORY COMPARISON

No. 0318-OV-A0312-18

Subject of the ILC:

GAUGE BLOCKS up to 100 mm

Carried out in the period of: April 2018 – November 2018

ILC identification: **11-0300-18**

Participant:

S.C. METROMAT S.R.L.

Str. Ady Endre nr. 44, 505600 Sacele, jud. Brasov, Romania

Final statement

Following the Report No. 0318-ZV-A0300-18 on evaluation of the ILC measurement results the above mentioned participant

c o m p l i e s w i t h

the requirements of this interlaboratory comparison based on EN ISO/IEC 17043.

Date of issue: **05.12.2018**

Český metrologický institut

-4-

RNDr. Simona Klenovská
Head of the ILC Department

Ing. František Staněk, PhD.
Director for Legal Metrology

12 laboratoare participante	Valori E_n obtinute de Metromat
Cehia (8); Israel(1); Bulgaria (1); Portugalia(1); Romania (1) (Metromat)	Cale plan paralele : E_n = - 0,22; 0,31; 0,12; -0.03; 0,39; 0,86; 0,25; 0,28 **Media valorilor E_n: 0,307**